Quolis Chair (Model Q5500) INSTALLATION INSTRUCTIONS



1 Risk Level Interpretation

Precautions before use

Make sure to carefully read the Safety Precautions and use the product correctly.

These precautions are intended to ensure the safe use of the product and prevent harm or damage to users or other people.

According to the magnitude of harm and damage and the degree of urgency, an incident that may be caused by misuse of the product is classified into one of the following categories: CONTRAINDICATION, WARNING, and CAUTION. All of these categories are important for safety. Always follow the instructions provided.

We assume no responsibility for any accident due to failure to follow the Safety Precautions even in the event of harm or damage to users or other persons. In such case, users or other persons who use the product without observing the Safety Precautions is responsible for any harm or damage incurred.

The graphical symbols are explained in detail below.

Once you have fully understood this explanation, read the text.

Classification by degree of harm or damage and urgency

CONTRAINDICATION

Use of the product without regard to this indication will create a hazardous condition that may result in death or serious injury.

MARNING

Improper handling of the product without regard to this indication will create a hazardous condition that may result in death or serious injury.

Improper handling of the product without regard to this indication will create a potentially hazardous condition that may result in moderate or slight injury or property damage.

The following graphical symbols are used to explain your responsibilities for using the product safely:

Graphical symbols for prohibited activity



Generally prohibited activity

Graphical symbol for mandatory instructions



Mandatory instructions for ordinary users

2 Safety Precautions

CONTRAINDICATION

N Installing or transferring the product

Precautions regarding installation

Do not install the product near electromagnetic sources such as communication facilities or elevators.

Malfunction of this product may occur in the presence of electromagnetic interference waves.

WARNING



Precautions for installation

Make sure to place the product on a firm and flat floor. Placing the equipment on a tilted floor may cause it to fall.

To avoid the risk of electric shock, this equipment must only be connected to a supply mains with protective earth.

Failure or electric leak may result in electric shock.

Installation

Precautions for operating the chair

Before and during the chair operation, pay close attention to the operation range (seat raising/lowering direction, backrest lowering/raising direction), and ensure that the chair does not contact any body parts, hands or feet, or objects. Body parts, hands or feet, or objects may become caught in the chair, causing injury or damage to peripheral devices.



Cautions on Installation Work

Metal or mechanical parts of this product may have sharp edges or corners. Be careful to avoid injury to your hands and fingers during installation work.

1 Guaranteeing safety of workers

- 1. The installation of this product and the connection of its supply connection part should be handled by a local authorized Belmont dealer.
- 2. Workers should wear work clothing that matches the operation, shoes (bare feet and sandals are strictly prohibited), and gloves for moving heavy objects.
- 3. Unless there are special reasons for operation, the supply of electricity, water, air and gas to the product should be shut down throughout operation.
- 4. To prevent the product from falling down, fix the chair on the floor before performing the assembly of unit parts, etc.

2 Precautions for installation

- 1. When moving the product, hold only the designated parts. Holding other sections may result in personal injury or damage to property.
- 2. Be sure to confirm that the retaining bolts are fixed before lifting up the chair.
- 3. Carry the equipment carefully in order to prevent its fall.
- 4. Do not let the product fall, crash into other objects, or apply strong shock to it.
- 5. Accurately undertake the rising process on the floor.
- 6. Be sure to remove dust in the source line before connecting the junction section.
- 7. Securely connect the water supply and drain hoses. Take deodorization measures for the drain hose.
- 8. Be sure to take fixation measures onto the floor.
- 9. Do not connect the power input to other than rated power supply.
- 10. Be sure to connect grounding.
- 11. Perform horizontal adjustment so that the table or light will not be dislocated.
- 12. After completing all processes, confirm that each section operates normally, and that there is no water leak or air leak, before putting on the cover.

3 Mobile communication devices

Mobile communication devices may affect electrical medical devices. Therefore, the use of mobile communication devices is prohibited in a medical office or hospital environment.

4 Repair

Never remodel this product. For its repair, use parts designated by Takara Belmont.

Conditions for installation site

1 Conditions for installation site

- 1. The floor surface should be even and horizontal.
- 2. The floor strength of the building should be \geq 4,900 N/m2.
- 3. As the floor material of the building, plywood with the thickness of 25 mm as recommended by Takara Belmont, or a composite material that combines plywood with the thickness of 15 mm and particle board with the thickness of 20 mm, should be used.

Alternatively, a floor material with the strength greater than the above should be used.

- 4. This product satisfies the EMC (electromagnetic compatibility) standards. However, it should not be used adjacent to, or stacked with, other equipment. If the product is to be used in proximity with other equipment, normal operation of the respective devices should be confirmed in the planned layout.
- 5. The product should be installed in a place that is not exposed to water.
- 6. The product should be installed in a place that will not be negatively affected by air pressure, temperature, humidity, ventilation, sunlight, dust, salt, sulfur content, and/or ozonic air (from an ozone generator).
- 7. Be careful about stability conditions, such as inclination, vibration and shock (including those during transport).
- 8. Do not install the product in a place where chemicals are stored or where gas is generated.
- 9. Be careful about the frequency and voltage of power supply, and about allowable ampacity, and connect grounding correctly.
- 10. To prevent water leaks, perform caulking securely in the penetrated section on the floor surface for riser piping.

2 Fixation on the floor

Concrete floor

Use anchor bolts with the diameter of 8 mm. The anchor bolts should be 35mm or more in depth.

Wood floor

Use coach bolts enclosed in the product package.

WARNING

Unless the this product is fixed onto the floor, there is risk of fall of the equipment, which may result in serious injury or damage not only to workers engaged in the installation process, but also to equipment in the clinic, users, patients, etc.

Conditions for installation site

- 3 Capacity of power supply equipment to be used
 - 1. The power supply specifications of the installation site should be observed. Rated voltage and power supply input for each Quolis Chair (Model Q5500) are following below.

The minimum capacity of power supply equipment should have 1,500 VA and over.

Rated voltage	AC120V
Power frquency	60Hz
Power supply input	4.2A

2. Be sure to connect grounding.

Technical Data

Operation mode

Chair Initial height/stroke

Chair raising and lowering

Weight

mechanism Headrest

Backrest

Armrest

Upholstered parts

Transportation/storage

Patient weight limit

Use in high-oxygen

Operation switch Usage environment

environment

environments

Seat

Model Classification for protection against electric shock Classification according to the degree of protection against electric shock Rated voltage Power frequency Power input Fuse AC-Q55-CD36* Class I Equipment

AC120 V

Type B Applied Parts(chair upholstered parts/armrest)

60 Hz 4.2 A Fuse holder 10A / 250V (Interrupting capacity: 750A / 250VAC) Operating speed: Interrupting Size: 6.3 × 32 mm Chair control board 1.25A / 250V (Interrupting capacity: 1500A / 250VAC) Operating speed: Interrupting Size: 5.2 × 20 mm Non-continuous operation (motor) Maximum operating time, 3 minutes DUTY 1:15 293 lbs (133 kg) 15.7 / 15.7 inch (400 / 400 mm) Electrohydraulic

Manual Electrohvdraulic Backrest-linked tilt mechanism With rotation function on both sides Synthetic leather Touchpad control panel/Foot switch Temperature 32 ~ 104°F (0 ~ 40°C) Humidity 10 ~ 95% (No condensation) Atmospheric pressure 10 ~ 15.4 psi (700 ~ 1060 hPa) Temperature: -4-158°F (-20-70°C) Humidity: 10-95% (No condensation) Atmospheric pressure: 10-15.4psi (700-1,060 hPa) 440 lbs (200 kg) The product is not suitable for use in a high-oxygen environment.

Refer to the rating plate for the capacity of power supply.

The specification is subject to change without notice.

Technical Data

19-1/8 (485) 8-5/8 (220)

26 (660)



Unit : inch (mm) Tolerance in dimensions: ±10%

The specification is subject to change without notice.

Installation

1 Necessary tools

No.	Tool name	QTY.
1	Phillips screwdriver No. 1–3	1 for each
2	6-piece double head open end wrench set (5.5–24 mm)	1
3	8-piece hex key wrench set (1.5–8 mm)	1

2 Supplied parts

Chair (unit)

No.	Part name	QTY.
1	Backrest unit	1
2	Backrest cover	1
3	Headrest cover	1
4	4 Armrest	
5	Hole plug	1
6	7Truss head screw, M5×10, plated108Countersunk screw, M5 × 8, plated39Pan head screw, 2P, M6×20, plated2	
7		
8		
9		
10		

Chair (cover parts)

No.	Part name	QTY.
1	Backrest cushion	1
2	Headrest cushion	1
3	Seat	1
4	Armrest cushion	2

Installation





STEP3 : Attaching the backrest and headrest
1) Attach the backrest plate assembly.
(2) Attach the backrest cover to the backrest plate assembly.
(3) Attach the backrest cushion to the backrest plate assembly. Insert the T-hooks through the 4 holes on the backrest cushion downward to lock.

④ Attach the headrest cushion to the sliding headrest mechanism with 3 screws, then attach the headrest cover.

STEP4 : Attaching the seat cushion and armrests

- (5) Insert the back of the seat cushion into the chair frame, and lock it in place with 2 screws from underneath the legrest area.
- (6) Attach the armrest cushions to the armrests, then insert the armrests into armrest brackets.

Concerning the dental unit to be attached to this chair, refer to the installation instructions for the unit.

2022-02	2022-02 (2r	
REF		1E08DUBE

(]Belmont

Importer in U.S.A. BELMONT EQUIPMENT, Division of Takara Belmont USA, Inc. 101 Belmont Drive, Somerset, New Jersey 08873 U.S.A. TEL.: (732) 469-5000 / (800) 223-1192 FAX.: (732) 356-1035

Importer in Canada TAKARA COMPANY, CANADA, LTD.

2455 Meadowvale Blvd. Mississauga, Ontario L5N 5S2, Canada TEL : (905) 816-8965 www.belmont.ca

BELMONT MANUFACTURING CO., LTD. (Manufacturer)

Lot I-2, Long Duc Industrial Park, Long Duc Ward, Long Thanh District, Dong Nai Province, Vietnam TEL.: +84-2513-201-100 / FAX.: +84-2513-201-096