



Notes :

- (1) Local regulations provide that licensed plumbers+ electricians shall install utilities.
- (2) Make sure all plumbing conforms to prevailing local codes.
- (3) Plumbing central vacuum to the utility center should be specified by the central vacuum supplier and terminated in utility center with 5/8" O.D. tube perpendicular to the floor.
- (4) Gravity drain has 1-1/2" nominal pipe protruding 1" from finished floor.
Note : Contractor to supply and place trap in line to conform with local codes.
- (5) Electrical contractor to supply quad electrical outlet box.

Standard Utility Center Floor Template

**EVOGUE 'B' Delivery Systems
and BDS Delivery Systems**
for X-Calibur Series Chairs

