Top of X-ray Horizontal Arm Extends Approximately 3-1/4” Above the Top Edge of the Wall Plate

17-13/16”(REF.)

11-11/16”

5-13/16”

℄

⌀1”
Hole for 4 conductor cable

⌀1”
Hole for AC 120V

12-1/16”

10-1/8”

10”

11-3/16”

NOTE: Level top and side of plate prior to securing to the wall

Drill (4) ⌀3/16” pilot holes

1”
(REF.)

℄

16”

15/16”

1-1/16”

Lower Edge of Plastic Cover

Lower Edge of Plastic Cover

Minimum Mounting Height from Floor to Bottom Edge of Plate: 34-1/2”

FLOOR

Bel-Ray II (Model 097)
PHOT-xIIs (Model 505)
Wall Mount Template

Belmont
1-800-223-1192

TEMPL015 Rev 102815